Standard Data Logger Gl







Order Reference


Standard Calibration Points

GMP-F-I P


• -80°C


• 140°C


• 0°C



- Recover the data through the docking station using XpertLog Software
- ITS-90 calibration coefficients are stored in the internal memory
- Complies with FDA 21 CFR Part 11
- Network ready: multiple users, centralized, server database, active directory authentication
- Calibration can be performed on site







Sensor type

- flexible
- customized
- Sensor size
- cross section

Operating Range

- body

- Sensor Accuracy

Resolution Memory capacity

Memory type Sampling rate

Internal clock drift Mechanical

- material

- enclosure
- weight
- diameter
- height Power

Battery lifetime

Docking station

Calibrations

Certification

4 wires PT-100 430 mm (17")

up to 5000 mm (195") 1.5 x 8 mm (0.06 x 0.315")

1.77 mm² (0.0027 inch²)

-90°C to +140°C -200°C to +200°C

±0.1°C from -90°C to 140°C

0.01°C

32,000 data points non volatile

1 sec to 24 hours

4 sec / 24 h @ 23°C

Peek 1000 - FDA compliant

IP-68

44 g (1.55 oz) 40.0 mm (1.57")

35.0 mm (1.38")

1 x Lithium batteries, 0.35 g/each Lithium content

classified UN3091 under IATA DGR section 3.9.2.6

compliant with ICAO/IATA packing instructions 970, user replaceable

Up to 12 months, based on usage USB to computer

4, 8, 12 and 20 ports available

Programs and reads unlimited number of loggers

Factory Calibrations: traceable NIST/COFRAC - ITS-90 coefficients stored

in the internal memory

User Calibrations: closed loop calibration using XpertLog® software

Health Safety: EN 60950-1 / 2006 EMC: EN 60601-1-2 / 2007









