




- ITS-90 calibration coefficients are stored in the internal memory
- Complies with FDA 21 CFR Part 11
- Network ready: multiple users, centralized, server database, active directory authentication
- Calibration can be performed on site
- Can be remotely activated







Sensor type - flexible

- customized

- Sensor size - cross section

Operating Range

- body

- Sensor

Accuracy Resolution

Memory capacity Memory type

Sampling rate Transmission

Internal clock drift Mechanical

> - material - enclosure

- weight - diameter

- height

Power

Battery lifetime Antenna Wireless range Radio frequency

Access point

Calibrations

Certification

4 wires

4 wires: 430 mm (17") up to 5000 mm (195") 1.5 x 8 mm (0.06 x 0.315") 1.77 mm² (0.0027 inch²)

-90°C to +140°C -200°C to +200°C

±0.1°C from -90°C to +140°C

0.01°C

32,000 data points non volatile 1 sec to 24 hours

simultaneus: maximum 20 loggers sequential: unlimited

4 sec / 24 h @ 23°C

Peek 1000 - FDA compliant

IP-68

50 g (1.76 oz) 37 mm (1.46")

62.81 mm (2.47")

2 x Lithium batteries, 0.35 g/each Lithium content classified UN3091 under IATA DGR section 3.9.2.6

compliant with ICAO/IATA packing instructions 970, user replaceable

Up to 6 months, based on usage Internal

121 m (400 feet)

2.4 Ghz ISM band USB to computer

Programs and reads unlimited number of loggers

Factory Calibrations: traceable NIST/COFRAC - ITS-90 coefficients stored

in the internal memory

User Calibrations: closed loop calibration using XpertLog® software

Health Safety: EN 60950-1 / 2006 EN 60601-1-2 / 2007 EMC:

EN 300440-2 v1.3.1 (2009-03)











Order Reference

GMP-RF-F-30H

GMP-RF-F-30H-2m GMP-RF-F-30H-3m

GMP-RF-F-30H-4m

GMP-RF-F-30H-5m

Standard Calibration Points

• -80°C

0°C

140°C

