






Temperature and Pressure Data Logger with Rigid Sensor 0.01 BarA to 5.0 BarA -40° C to $+140^{\circ}$ C



- Collects and stores pressure and temperature information in the internal memory
- Recover the data through the docking station using XpertLog Software
- ITS-90 calibration coefficients are stored in the internal memory
- Complies with FDA 21 CFR Part 11
- Network ready: multiple users, centralized, server database, active directory authentication







Sensor type

- rigid

- Sensor size

- cross section Operating Range

- body

- Sensor

Accuracy

Resolution Memory capacity

Memory type Sampling rate Internal clock drift

Mechanical

- material - enclosure

- weight

- diameter

- height Power

Battery lifetime

Docking station

Calibrations

Certification

Absolute pressure / Temp: 4 wires PT-100

20 mm (0.8")

1.5 x 8 mm (0.06 x 0.315") 1.77 mm² (0.0027 inch²)

0.01 BarA to 5.0 BarA -40°C to +140°C

-40°C to +140°C

±10 mbar ≤ 3 BarA @ 130°C; ±15 mbar > 3 BarA @ 130°C; ±15 mbar \leq 3 BarA @ 23°C; \pm 30 mbar > 3 BarA @ 23°C; \pm 0.1°C from -40°C to

+140°C 0.1 mbar; 0.01°C

16,000 data points pressure / 16,000 data points temperature

non volatile 1 sec to 24 hours 4 sec / 24 h @ 23°C

Peek 1000 - FDA compliant

IP-51

57 g (2.01 oz) 40.0 mm (1.57")

69.0 mm (2.72")

1 x Lithium batteries, 0.35 g/each Lithium content

classified UN3091 under IATA DGR section 3.9.2.6 compliant with ICAO/IATA packing instructions 970, user replaceable

Up to 12 months, based on usage

USB to computer

4, 8, 12 and 20 ports available

Programs and reads unlimited number of loggers

Factory Calibrations: traceable NIST/COFRAC - ITS-90 coefficients stored

in the internal memory

User Calibrations: closed loop calibration using XpertLog® software

EN 60950-1 / 2006 Health Safety: EMC: EN 60601-1-2 / 2007



Order Reference

GMP-TP

Standard Calibration Points

hysteresis at 23°C and 130°C:

- 0.05 BarA
- 0.50 BarA
- 1.00 BarA
- 1.50 BarA
- 2.00 BarA
- 2.50 BarA
- 3.00 BarA
- 3.50 BarA
- 4.00 BarA
- 4.50 BarA
- 5.00 BarA









